Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



Accredited to ISO/IEC 17025:2005

Furness Controls Limited

Issue No: 031 Issue date: 08 April 2016

 Beeching Road
 Contact: Mr D Walker

 Bexhill
 Tel: +44 (0)1424-730316

 East Sussex
 Fax: +44 (0)1424-730317

TN39 3LJ E-Mail: calibration@furness-controls.com
Website: www.furness-controls.com

Calibration performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address Beeching Road Bexhill East Sussex TN39 3LJ	Local contact Doug Walker	Flow calibration Pressure calibration Electrical Calibration	Perm
Techniparc 3 rue Boole 91240 St. Michel sur Orge France	Jean-Philippe Noblet Tel.+33 1 69460020	Flow calibration Pressure calibration Electrical Calibration	France and site
Karl Arnold Strasse 12 D-47877 Willich Munchheide (2) Germany	Karsten Bartsch Tel. +49 21 54 49 96 80	Flow calibration Pressure calibration Electrical Calibration	Germany and site

Site activities performed away from the locations listed above:

Location details		Activity	Location code
The customer's site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer	Site contact: Sarah Hedge 4 The Pavilions Amber Close Tamworth Staffordshire B77 4RP Tel: +44 (0)1827 59950	Flow calibration Pressure calibration	Site

Assessment Manager: JG2 Page 1 of 4



Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Furness Controls Limited

Issue No: 031 Issue date: 08 April 2016

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k = 2)	Remarks	Location Code
PRESSURE				
Gas pressure (gauge)				
Calibration of pressure indicating instruments and gauges Pressure equivalent calibration of Furness controls FRS 4's and other dead weight testers	-100 kPa to -10 kPa -10 kPa to 0 Pa 0 Pa to 3 kPa 3 kPa to 12 kPa 12 kPa to 30 kPa 30 kPa to 40 kPa 40 kPa to 4 MPa	0.010 % 0.010 % + 0.30 Pa 0.010 % + 0.030 Pa 0.010 % + 0.30 Pa 0.010 % + 1.0 Pa 0.010 % + 2.0 Pa 0.010 %	Calibrations of devices with an electrical output may be undertaken	Perm
Gas pressure (absolute)				
Calibration of pressure indicating instruments and gauges Gas pressure (gauge)	0 Pa to 4 MPa	0.010 % + 10 Pa		Perm
Calibration of pressure indicating instruments and gauges	-100 kPa to 0 Pa 0 Pa to 220 Pa 220 Pa to 2200 Pa 2.2 kPa to 22 kPa 22 kPa to 100 kPa 100 kPa to 400 kPa 400 kPa to 1.6 MPa 1.6 MPa to 4 MPa	0.25 % + 100 Pa 0.30 % + 0.10 Pa 0.30 % + 0.60 Pa 0.30 % + 6.0 Pa 0.25 % + 50 Pa 0.25 % + 200 Pa 0.25 % + 800 Pa 0.25 % + 2.0 kPa	Calibrations of devices with an electrical output may be undertaken	Site
Gas pressure (absolute)				
Calibration of pressure indicating instruments and gauges	0 Pa to 160 kPa	0.25 % + 100 Pa		Site
FLOW				
Flow Rate - Gas Volume	0.02 ml/min to 2ml/min 2 ml/min to 2000 l/min	0.75 % 0.60 %	Calibration medium Air. Calibrations up to 10 l/min can be undertaken on Nitrogen.	Perm
Gas - Volume Passed	180 I to 300 I (at flow rates of 2 I/min to 2000 I/min)	0.36 %		Perm
	0.02 ml to 6000 l (at flow rates of 0.02 ml/min to 2ml/min 2 ml/min to 2000 l/min)	0.75 % 0.50 %		Perm

Assessment Manager: JG2 Page 2 of 4



Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Furness Controls Limited

Issue No: 031 Issue date: 08 April 2016

Calibration performed by the Organisation at the locations specified

Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k = 2)	Remarks	Location Code
0.04 ml/min to 0.4 ml/min 0.4 ml/min to 40 ml/min 40 ml/min to 2000 l/min	1.3 % + 0.0020 ml/min 1.3 % 1.2 %	Calibration medium Air. Calibrations of devices with an electrical output may be undertaken	Site
0 V to 10 V 10 V to 55 V	30 ppm + 20 μV 30 ppm + 200 μV		Perm France Germany
0 A to 10 mA 10 mA to 100 mA	100 ppm + 0.30 μA 100 ppm + 3.0 μA		Perm France Germany
		Calibrations of devices with an electrical output may be undertaken	France
0.04 ml/min to 0.4 ml/min 0.4 ml/min to 40 ml/min 40 ml/min to 400 l/min	1.1 % + 0.0010 ml/min 1.1 % 1.0 %	Calibration medium Air	
-100 kPa to -10 kPa -10 kPa to 0 Pa 0 Pa to 3 kPa 3 kPa to 12 kPa 12 kPa to 30 kPa 30 kPa to 40 kPa 40 kPa to 100 kPa 100 kPa to 1.6 MPa 1.6 MPa to 4 MPa	0.10 % + 100 Pa 0.014 % + 0.50 Pa 0.014 % + 0.040 Pa 0.014 % + 0.50 Pa 0.014 % + 1.0 Pa 0.014 % + 2.0 Pa 0.10 % + 20 Pa 0.10 % + 150 Pa 0.10 % + 200 Pa		
1 kPa to 160 kPa 160 kPa to 200 kPa	0.10 % + 100 Pa 0.10 % + 200 Pa		
	0.04 ml/min to 0.4 ml/min 0.4 ml/min to 40 ml/min to 2000 l/min 0 V to 10 V 10 V 10 V to 55 V 0 A to 10 mA 10 mA to 100 mA 0.04 ml/min to 40 ml/min 40 ml/min to 400 l/min -100 kPa to -10 kPa -10 kPa to 0 Pa 0 Pa to 3 kPa 3 kPa to 12 kPa 12 kPa to 30 kPa 30 kPa to 100 kPa 40 kPa to 1.6 MPa 1.6 MPa to 4 MPa 1.6 MPa to 160 kPa	Range Range Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k = 2) 0.04 ml/min to 0.4 ml/min 0.4 ml/min to 40 ml/min 40 ml/min to 2000 l/min 1.3 % + 0.0020 ml/min 1.3 % 1.2 % 0 V to 10 V 10 V to 55 V 30 ppm + 20 μV 30 ppm + 200 μV 0 A to 10 mA 100 ppm + 0.30 μA 100 ppm + 3.0 μA 100 ppm + 0.30 μA 100 ppm + 0.4 ml/min 1.1 % 1.0 % 1.1 % + 0.0010 ml/min 1.1 % 1.0 % 0.014 % + 0.50 Pa 0.014 % + 2.0 Pa 0.10 % + 20 Pa 0.10 % + 150 Pa 0.10 % + 200 Pa	Range Range Remarks Remarks Remarks Remarks

Assessment Manager: JG2 Page 3 of 4



Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Furness Controls Limited

Issue No: 031 Issue date: 08 April 2016

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k = 2)	Remarks	Location Code
German Capability			Calibrations of devices with an electrical output may be undertaken	Germany
FLOW				
Flow Rate - Gas Volume	0.04 ml/min to 0.4 ml/min 0.4 ml/min to 40 ml/min 40 ml/min to 400 l/min	1.1 % + 0.0010 ml/min 1.1 % 1.0 %	Calibration medium Air	
PRESSURE				Germany
Gas pressure (gauge)				
Calibration of pressure indicating instruments and gauges Furness controls FRS 4's. pressure equivalent	-100 kPa to -10 kPa -10 kPa to 0 Pa 0 Pa to 3 kPa 3 kPa to 12 kPa 12 kPa to 30 kPa 30 kPa to 40 kPa 40 kPa to 400 kPa 400 kPa to 1.6 MPa	0.15 % + 100 Pa 0.014 % + 0.50 Pa 0.014 % + 0.040 Pa 0.014 % + 0.50 Pa 0.014 % + 1.0 Pa 0.014 % + 2.0 Pa 0.15 % + 50 Pa 0.15 % + 150 Pa		
Gas pressure (absolute)				
Calibration of pressure indicating instruments and gauges	1 kPa to 120 kPa	0.15 % + 100 Pa		
END				

Page 4 of 4 Assessment Manager: JG2